

AMENDMENTS TO THE CLAIMS:

Please cancel claims 4 and 11 without prejudice or disclaimer, and amend claims 1-3, 5, 8-10 and 12-14, as follows. This listing of claims will replace all prior versions, and listings, of claims in the application:

Listing of Claims:

Claim 1 (Currently amended): An embolus forming in-vivo indwelling device comprising a coil separating member and a coil main body having flexibility and a stretch suppressing member which is provided on an inner periphery of the coil main body and which is made of a water-swellaable polymer material for suppressing stretch of the coil main body by swelling with absorbed water,

wherein the stretch suppressing member has a rod-like shape and is provided in the coil main body so as to pass through the coil main body and extend in the coil axial direction of the coil main body,

wherein in a case that the [[dry]] stretch suppressing member is provided on the inner periphery of the coil main body, the stretch suppressing member, when dry, has a smaller diameter than the coil diameter of the coil main body, and the stretch suppressing member enters space between adjacent wire turns of the coil main body as a result of swelling.

Claim 2 (Currently amended): The embolus forming in-vivo indwelling [[coil]] device according to claim 1, wherein the water-swellaable polymer material constituting the stretch suppressing member comprises a polyvinyl alcohol polymer.

Claim 3 (Currently amended): The embolus forming in-vivo indwelling [[coil]] device according to claim 1, wherein the ~~wire constituting the coil main body has a diameter of 10 to 120 μ m, and the coil main body~~ is constituted of a wire that has a coil diameter of 100 to 500 μ m, a coil length of 2 to 500 mm, [[and]] a number of turns of 1 to 100 per unit length (1 mm), and a diameter of 10 to 120 μ m.

Claim 4 (Canceled).

Claim 5 (Currently amended): The embolus forming in-vivo indwelling [[coil]] device according to claim 4, wherein the diameter of the stretch suppressing member is smaller than the inner diameter of the coil main body by about 1 to 50% in a dry state.

Claims 6-7 (Canceled).

Claim 8 (Withdrawn): The embolus forming in-vivo indwelling [[coil]] device according to claim 1, wherein a stretch suppressing member has a rod-like or cylindrical shape and is provided so as to extend in the coil axial direction of a coil main body and pass through the coil main body, and another stretch suppressing member has a cylindrical or tubular shape and is provided to cover the entire region of the outer periphery of the coil main body in the coil axial direction.

Claim 9 (Withdrawn): The embolus forming in-vivo indwelling [[coil]] device according to claim [[4]] 1, further comprising another stretch suppressing member having cylindrical or tubular shape and is provided to cover the entire region of the outer periphery of the coil main body in the coil axial direction.

Claim 10 (Withdrawn): The embolus forming in-vivo indwelling [[coil]] device according to claim 1, wherein the stretch suppressing member extends over the entire region of the coil main body.

Claim 11 (Canceled).

Claim 12 (Currently amended): The embolus forming in-vivo indwelling [[coil]] device according to claim 1, wherein the coil main body has space between each turn before swelling.

Claim 13 (Withdrawn): The embolus forming in-vivo indwelling [[coil]] device according to claim 1, wherein the stretch suppressing member is disposed to extend over the entire region of the coil main body in the coil axial direction.

Claim 14 (Currently amended): The embolus forming in-vivo indwelling [[coil]] device according to claim 1, wherein [[a]] the coil separating member is partially provided inside

U.S. Patent Application Serial No.: **10/541,469**

Amendment filed April 14, 2011

Reply to OA dated January 19, 2011

the coil body in a state that a distal end of the coil separating member is in contact with a proximal end of the stretch suppressing member.